

Epalinges: Xenogen system back online

Contributed by Yannick KREMPP
Wednesday, 30 March 2016
Last Updated Thursday, 31 March 2016

Dear CIF users,

The Xenogen system located at the Epalinges campus is out of order until further notice.

We will update the status of the device here as soon as we have news to share, so please come back to check this page later.

Status update:

30.03.2016 : System down. Full check-up scheduled for tomorrow.

31.03.2016 : The check-up is planned for this afternoon at 13h30. We will let you know the status of the instrument afterwards.

Some users have asked about their bookings: they will off course be removed and you will not be charged for not using a device that malfunctioned

31.03.2016: (14h30) The system is up and running again. The camera cable has been slightly dislodged from its port, causing an error during the initialization.

Previous bookings have been removed.